

Vaporiser Calibration Report

Overview

Vaporiser Serial, Type	BEJK 12345 T# 00001, Tec 7 SEVOFLURANE	
Customer	ABC Hospital	
Environment	21.00°C, 101353 Pa	
Operator	Mary Smith	
Carrier A	Oxygen	100.00%
Carrier B	(none)	0.00%
Agent	Sevoflurane	0.00%

Results

Setting	5.00 lpm
0.00%	0.00
0.60%	0.55
1.00%	1.01
2.00%	2.01
3.00%	3.02
4.00%	4.09
5.00%	5.02
6.00%	5.95
7.00%	7.14
8.00%	8.01

This vaporiser has been serviced, calibrated, and certified to be within the correct operating criteria.

QUALITY CONTROL
FINAL INSPECTION
DATE <i>N/A</i>
INITIAL

FINAL LEAK TEST
PASS
DATE <i>N/A</i>
INITIAL